

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070725\
 Data File : VX046902.D
 Acq On : 07 Jul 2025 13:07
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 08 04:52:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 08 04:37:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	387345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	679390	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	616906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	306943	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	261983	51.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.400%
35) Dibromofluoromethane	5.397	113	222695	48.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	813973	50.026	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.060%
62) 4-Bromofluorobenzene	11.079	95	306274	49.527	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070725\
Data File : VX046902.D
Acq On : 07 Jul 2025 13:07
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 08 04:52:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 08 04:37:37 2025
Response via : Initial Calibration

