

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010696.D
 Acq On : 08 Jul 2019 15:26
 Operator : JC/SP
 Sample : K3669-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW218-190702

Quant Time: Jul 09 01:53:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	226572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	354960	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	324938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150425	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113820	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
35) Dibromofluoromethane	5.50	113	109551	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	8.71	98	419561	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.13	95	141987	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
Target Compounds						
3) Chloromethane	1.33	50	2104	1.147	ug/l	91
4) Vinyl Chloride	1.41	62	8180	4.072	ug/l	94
16) Acetone	2.44	43	8410	7.899	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	8864	3.632	ug/l	91
40) Benzene	6.13	78	9449	1.075	ug/l #	89
65) Chlorobenzene	10.14	112	58674	8.894	ug/l	100
83) tert-Butylbenzene	11.77	119	3519	0.417	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010696.D
Acq On : 08 Jul 2019 15:26
Operator : JC/SP
Sample : K3669-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW218-190702

Quant Time: Jul 09 01:53:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

