

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010706.D
 Acq On : 08 Jul 2019 19:19
 Operator : JC/SP
 Sample : K3664-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0582

Manual Integrations
 APPROVED

MMDadoda
 7/9/2019 11:12:24 AM

Quant Time: Jul 09 02:29:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	370537	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	344356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176291	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117941	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
35) Dibromofluoromethane	5.50	113	112630	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
50) Toluene-d8	8.71	98	448168	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.13	95	159797	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	5655	4.962	ug/l	95
31) Cyclohexane	5.59	56	680942	207.327	ug/l	98
39) Methylcyclohexane	7.46	83	1065917	275.399	ug/l	97
40) Benzene	6.15	78	621116	67.708	ug/l	99
67) Ethyl Benzene	10.25	91	2915810	247.051	ug/l	100
68) m/p-Xylenes	10.35	106	443000	96.869	ug/l	99
69) o-Xylene	10.69	106	3716	0.826	ug/l	98
73) Isopropylbenzene	11.02	105	940622	78.711	ug/l	99
78) n-propylbenzene	11.36	91	1540382	117.282	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	297678	30.216	ug/l	98
83) tert-Butylbenzene	11.77	119	17769m	1.797	ug/l	
84) 1,2,4-Trimethylbenzene	11.81	105	5090798	514.336	ug/l	99
85) sec-Butylbenzene	11.94	105	206961	17.826	ug/l	85
86) p-Isopropyltoluene	12.06	119	81895	7.626	ug/l	98
89) n-Butylbenzene	12.38	91	98082m	10.748	ug/l	
95) Naphthalene	13.83	128	1071606	86.860	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010706.D
Acq On : 08 Jul 2019 19:19
Operator : JC/SP
Sample : K3664-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0582

Manual Integrations
APPROVED

MMDadoda
7/9/2019 11:12:24 AM

Quant Time: Jul 09 02:29:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

