

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010715.D
 Acq On : 08 Jul 2019 23:36
 Operator : JC/SP
 Sample : K3691-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E3-TB01-190704

Quant Time: Jul 09 03:33:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230332	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	359810	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	325991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149983	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115456	53.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.92%	
35) Dibromofluoromethane	5.50	113	109600	49.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.58%	
50) Toluene-d8	8.71	98	423381	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
62) 4-Bromofluorobenzene	11.13	95	144917	47.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010715.D
Acq On : 08 Jul 2019 23:36
Operator : JC/SP
Sample : K3691-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E3-TB01-190704

Quant Time: Jul 09 03:33:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

