

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010725.D
 Acq On : 09 Jul 2019 03:30
 Operator : JC/SP
 Sample : K3690-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS038MW103-190703

Quant Time: Jul 09 09:29:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	230618	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	362541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167288	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117018	54.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.24%	
35) Dibromofluoromethane	5.50	113	109449	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
50) Toluene-d8	8.71	98	438894	51.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.13	95	152946	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3540	3.267	ug/l #	78
31) Cyclohexane	5.58	56	1600	0.512	ug/l	86
40) Benzene	6.14	78	4115	0.458	ug/l	92
65) Chlorobenzene	10.14	112	534841	78.004	ug/l	98
73) Isopropylbenzene	11.02	105	13956	1.231	ug/l	98
83) tert-Butylbenzene	11.77	119	68409	7.291	ug/l	91
85) sec-Butylbenzene	11.94	105	100788	9.149	ug/l #	55
88) 1,4-Dichlorobenzene	12.10	146	9014	1.602	ug/l #	78
91) 1,2-Dichlorobenzene	12.39	146	4268	0.777	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010725.D
Acq On : 09 Jul 2019 03:30
Operator : JC/SP
Sample : K3690-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS038MW103-190703

Quant Time: Jul 09 09:29:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

