

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010729.D
 Acq On : 09 Jul 2019 05:03
 Operator : JC/SP
 Sample : K3691-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0535-190704

Quant Time: Jul 09 09:55:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	227588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	354112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	321666	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	154625	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113388	53.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.28%	
35) Dibromofluoromethane	5.50	113	106560	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	8.71	98	420666	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
62) 4-Bromofluorobenzene	11.13	95	145417	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
Target Compounds						
4) Vinyl Chloride	1.41	62	9398	4.657	ug/l	94
65) Chlorobenzene	10.14	112	37083	5.678	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010729.D
Acq On : 09 Jul 2019 05:03
Operator : JC/SP
Sample : K3691-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0535-190704

Quant Time: Jul 09 09:55:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

