

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
 Data File : VX010685.D
 Acq On : 08 Jul 2019 10:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 01:20:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.348	0.422	-21.3#	107	0.00
3 P	Chloromethane	0.405	0.422	-4.2	92	0.00
4 C	Vinyl Chloride	0.443	0.455	-2.7#	92	0.00
5 T	Bromomethane	0.216	0.229	-6.0	96	0.00
6 T	Chloroethane	0.262	0.295	-12.6	99	0.00
7 T	Trichlorofluoromethane	0.703	0.779	-10.8	96	0.00
8 T	Diethyl Ether	0.252	0.262	-4.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.426	0.454	-6.6	94	0.00
10 T	Methyl Iodide	0.608	0.584	3.9	78	0.00
11 T	Tert butyl alcohol	0.123	0.105	14.6	76	0.00
12 CM	1,1-Dichloroethene	0.433	0.432	0.2#	86	0.00
13 T	Acrolein	0.081	0.080	1.2	89	0.00
14 T	Allyl chloride	0.615	0.647	-5.2	89	0.00
15 T	Acrylonitrile	0.243	0.240	1.2	84	0.00
16 T	Acetone	0.235	0.288	-22.6#	99	0.00
17 T	Carbon Disulfide	1.163	1.174	-0.9	88	0.00
18 T	Methyl Acetate	0.468	0.466	0.4	86	0.00
19 T	Methyl tert-butyl Ether	1.363	1.399	-2.6	89	0.00
20 T	Methylene Chloride	0.484	0.495	-2.3	90	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.469	-0.6	88	0.00
22 T	Diisopropyl ether	1.286	1.329	-3.3	90	0.00
23 T	Vinyl Acetate	1.111	1.199	-7.9	90	0.00
24 P	1,1-Dichloroethane	0.744	0.768	-3.2	89	0.00
25 T	2-Butanone	0.345	0.363	-5.2	87	0.00
26 T	2,2-Dichloropropane	0.646	0.685	-6.0	92	0.00
27 T	cis-1,2-Dichloroethene	0.539	0.537	0.4	86	0.00
28 T	Bromochloromethane	0.356	0.341	4.2	84	0.00
29 T	Tetrahydrofuran	0.213	0.208	2.3	83	0.00
30 C	Chloroform	0.800	0.824	-3.0#	90	0.00
31 T	Cyclohexane	0.677	0.713	-5.3	91	0.00
32 T	1,1,1-Trichloroethane	0.706	0.743	-5.2	90	0.00
33 S	1,2-Dichloroethane-d4	0.469	0.454	3.2	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.306	0.296	3.3	78	0.00
36 T	1,1-Dichloropropene	0.397	0.423	-6.5	90	0.00
37 T	Ethyl Acetate	0.409	0.429	-4.9	84	-0.01
38 T	Carbon Tetrachloride	0.430	0.467	-8.6	89	0.00
39 T	Methylcyclohexane	0.522	0.564	-8.0	91	0.00
40 TM	Benzene	1.238	1.282	-3.6	86	0.00
41 T	Methacrylonitrile	0.224	0.233	-4.0	86	0.00
42 TM	1,2-Dichloroethane	0.377	0.430	-14.1	93	0.00
43 T	Isopropyl Acetate	0.676	0.719	-6.4	87	0.00
44 TM	Trichloroethene	0.370	0.384	-3.8	86	0.00
45 C	1,2-Dichloropropane	0.313	0.321	-2.6#	86	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX070819\
 Data File : VX010685.D
 Acq On : 08 Jul 2019 10:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 01:20:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.224	0.236	-5.4	87	0.00
47 T	Bromodichloromethane	0.411	0.440	-7.1	87	0.00
48 T	Methyl methacrylate	0.314	0.339	-8.0	87	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	81	0.00
50 S	Toluene-d8	1.170	1.098	6.2	76	0.00
51 T	4-Methyl-2-Pentanone	0.433	0.459	-6.0	86	0.00
52 CM	Toluene	0.815	0.851	-4.4#	86	0.00
53 T	t-1,3-Dichloropropene	0.461	0.497	-7.8	88	0.00
54 T	cis-1,3-Dichloropropene	0.501	0.536	-7.0	86	0.00
55 T	1,1,2-Trichloroethane	0.336	0.349	-3.9	85	0.00
56 T	Ethyl methacrylate	0.509	0.531	-4.3	84	0.00
57 T	1,3-Dichloropropane	0.522	0.553	-5.9	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.234	3.3	77	0.00
59 T	2-Hexanone	0.340	0.371	-9.1	88	0.00
60 T	Dibromochloromethane	0.376	0.399	-6.1	86	0.00
61 T	1,2-Dibromoethane	0.363	0.387	-6.6	88	0.00
62 S	4-Bromofluorobenzene	0.423	0.418	1.2	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.404	0.406	-0.5	86	0.00
65 PM	Chlorobenzene	1.015	1.043	-2.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.390	-5.4	85	0.00
67 C	Ethyl Benzene	1.714	1.815	-5.9#	88	0.00
68 T	m/p-Xylenes	0.664	0.710	-6.9	88	0.00
69 T	o-Xylene	0.653	0.679	-4.0	87	0.00
70 T	Styrene	1.095	1.181	-7.9	88	0.00
71 P	Bromoform	0.322	0.339	-5.3	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.389	3.406	-0.5	89	0.00
74 T	N-amyl acetate	1.281	1.340	-4.6	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.122	1.086	3.2	85	0.00
76 T	1,2,3-Trichloropropane	0.911	0.930	-2.1	89	0.00
77 T	Bromobenzene	0.946	0.929	1.8	88	0.00
78 T	n-propylbenzene	3.725	3.921	-5.3	91	0.00
79 T	2-Chlorotoluene	2.218	2.255	-1.7	90	0.00
80 T	1,3,5-Trimethylbenzene	2.794	2.907	-4.0	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.391	0.371	5.1	81	0.00
82 T	4-Chlorotoluene	2.524	2.653	-5.1	91	0.00
83 T	tert-Butylbenzene	2.804	2.892	-3.1	90	0.00
84 T	1,2,4-Trimethylbenzene	2.807	2.957	-5.3	90	0.00
85 T	sec-Butylbenzene	3.293	3.419	-3.8	90	0.00
86 T	p-Isopropyltoluene	3.046	3.208	-5.3	91	0.00
87 T	1,3-Dichlorobenzene	1.652	1.683	-1.9	91	0.00
88 T	1,4-Dichlorobenzene	1.682	1.674	0.5	90	0.00
89 T	n-Butylbenzene	2.588	2.808	-8.5	93	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070819\
Data File : VX010685.D
Acq On : 08 Jul 2019 10:37
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 01:20:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.527	0.545	-3.4	86	0.00
91 T	1,2-Dichlorobenzene	1.642	1.622	1.2	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.237	2.1	89	0.00
93 T	1,2,4-Trichlorobenzene	1.139	1.161	-1.9	89	0.00
94 T	Hexachlorobutadiene	0.556	0.587	-5.6	92	0.00
95 T	Naphthalene	3.499	3.419	2.3	84	0.00
96 T	1,2,3-Trichlorobenzene	1.132	1.146	-1.2	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6