

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\
 Data File : VX017254.D
 Acq On : 08 Jul 2020 14:45
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 09 03:19:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	243956	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	415119	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	451108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	161412	53.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.34%	
35) Dibromofluoromethane	5.46	113	140134	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
50) Toluene-d8	8.70	98	523159	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
62) 4-Bromofluorobenzene	11.12	95	213654	56.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.58%	
Target Compounds						
16) Acetone	2.42	43	19771	17.573	ug/l	99
18) Methyl Acetate	2.75	43	3989	1.358	ug/l	98
20) Methylene Chloride	2.84	84	17142	5.954	ug/l	96
43) Isopropyl Acetate	6.42	43	74766	11.303	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\
Data File : VX017254.D
Acq On : 08 Jul 2020 14:45
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 09 03:19:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

