

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\
 Data File : VX017255.D
 Acq On : 08 Jul 2020 15:07
 Operator : JC/SP
 Sample : L3207-01 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31063

Quant Time: Jul 08 15:56:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	245781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	416819	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	428616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	222058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	164563	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
35) Dibromofluoromethane	5.47	113	142023	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
50) Toluene-d8	8.70	98	508390	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
62) 4-Bromofluorobenzene	11.12	95	207710	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
Target Compounds						
16) Acetone	2.42	43	11588	10.223	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\
Data File : VX017255.D
Acq On : 08 Jul 2020 15:07
Operator : JC/SP
Sample : L3207-01 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31063

Quant Time: Jul 08 15:56:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

