

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX070820\
 Data File : VX017282.D
 Acq On : 09 Jul 2020 01:41
 Operator : JC/SP
 Sample : L3235-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WGGM-20-WW-TSF-023

Quant Time: Jul 09 05:36:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X062520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jun 26 09:03:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	49241	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	251805	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	219023	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	107141	27.95	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	93.17%
60) 4-Bromofluorobenzene	11.12	95	107175	29.34	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	97.80%
63) Toluene-d8	8.70	98	247992	25.69	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	85.63%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX070820\
Data File : VX017282.D
Acq On : 09 Jul 2020 01:41
Operator : JC/SP
Sample : L3235-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WGGM-20-WW-TSF-023

Quant Time: Jul 09 05:36:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\624X062520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jun 26 09:03:25 2020
Response via : Initial Calibration

