

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\
 Data File : VX017285.D
 Acq On : 09 Jul 2020 02:49
 Operator : JC/SP
 Sample : PB130015TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130015TB

Quant Time: Jul 09 05:53:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	250095	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	422645	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	455225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	232524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	164913	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
35) Dibromofluoromethane	5.46	113	140436	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
50) Toluene-d8	8.70	98	519080	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
62) 4-Bromofluorobenzene	11.12	95	217033	56.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.32%	
Target Compounds						
16) Acetone	2.42	43	15444	13.390	ug/l	95
20) Methylene Chloride	2.84	84	16057	5.372	ug/l	97
43) Isopropyl Acetate	6.42	43	69847	10.371	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\
Data File : VX017285.D
Acq On : 09 Jul 2020 02:49
Operator : JC/SP
Sample : PB130015TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130015TB

Quant Time: Jul 09 05:53:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

