

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\
 Data File : VX017290.D
 Acq On : 09 Jul 2020 04:43
 Operator : JC/SP
 Sample : PB130015ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130015ZHE#05

Quant Time: Jul 09 06:22:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	247290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	428715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	462686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	243124	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166208	54.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.04%	
35) Dibromofluoromethane	5.46	113	142463	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
50) Toluene-d8	8.70	98	539410	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
62) 4-Bromofluorobenzene	11.12	95	223504	57.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.02%	
Target Compounds						
16) Acetone	2.42	43	17212	15.092	ug/l	93
18) Methyl Acetate	2.75	43	3539	1.189	ug/l	99
20) Methylene Chloride	2.84	84	20527	7.178	ug/l	96
43) Isopropyl Acetate	6.41	43	67828	9.929	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\
Data File : VX017290.D
Acq On : 09 Jul 2020 04:43
Operator : JC/SP
Sample : PB130015ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130015ZHE#05

Quant Time: Jul 09 06:22:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

