

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\  
 Data File : VX017291.D  
 Acq On : 09 Jul 2020 05:06  
 Operator : JC/SP  
 Sample : PB130015ZHE#06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 50 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB130015ZHE#06

Quant Time: Jul 09 06:24:01 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 29 17:10:38 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 5.63  | 168  | 248893   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 418259   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 455376   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 236433   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 163086   | 53.15  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 106.30% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 140919   | 53.26  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 106.52% |          |
| 50) Toluene-d8              | 8.70  | 98   | 531900   | 54.37  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 108.74% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 220797   | 57.73  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 115.46% |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 2.42  | 43   | 16380    | 14.270 | ug/l    | 94       |
| 18) Methyl Acetate          | 2.75  | 43   | 3544     | 1.183  | ug/l    | 100      |
| 20) Methylene Chloride      | 2.84  | 84   | 19639    | 6.784  | ug/l    | 96       |
| 43) Isopropyl Acetate       | 6.42  | 43   | 71303    | 10.698 | ug/l    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\  
Data File : VX017291.D  
Acq On : 09 Jul 2020 05:06  
Operator : JC/SP  
Sample : PB130015ZHE#06  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 50 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB130015ZHE#06

Quant Time: Jul 09 06:24:01 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 29 17:10:38 2020  
Response via : Initial Calibration

