

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070820\
 Data File : VX017273.D
 Acq On : 08 Jul 2020 22:16
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 05:39:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.484	0.451	6.8	90	0.00
3 P	Chloromethane	0.586	0.475	18.9	80	0.00
4 C	Vinyl Chloride	0.543	0.488	10.1#	88	0.00
5 T	Bromomethane	0.287	0.270	5.9	91	0.00
6 T	Chloroethane	0.321	0.316	1.6	98	0.00
7 T	Trichlorofluoromethane	0.803	0.869	-8.2	107	0.00
8 T	Diethyl Ether	0.293	0.300	-2.4	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.482	0.492	-2.1	98	0.00
10 T	Methyl Iodide	0.621	0.560	9.8	81	0.00
11 T	Tert butyl alcohol	0.118	0.128	-8.5	103	0.00
12 CM	1,1-Dichloroethene	0.495	0.477	3.6#	94	0.00
13 T	Acrolein	0.070	0.085	-21.4#	118	0.00
14 T	Allyl chloride	0.914	0.852	6.8	92	0.00
15 T	Acrylonitrile	0.283	0.302	-6.7	102	0.00
16 T	Acetone	0.231	0.253	-9.5	107	0.00
17 T	Carbon Disulfide	1.515	1.249	17.6	83	0.00
18 T	Methyl Acetate	0.602	0.682	-13.3	109	0.00
19 T	Methyl tert-butyl Ether	1.660	1.771	-6.7	101	0.00
20 T	Methylene Chloride	0.589	0.559	5.1	95	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.527	4.0	93	0.00
22 T	Diisopropyl ether	1.692	1.715	-1.4	95	0.00
23 T	Vinyl Acetate	1.454	1.517	-4.3	95	0.00
24 P	1,1-Dichloroethane	0.963	0.971	-0.8	97	0.00
25 T	2-Butanone	0.384	0.411	-7.0	100	0.00
26 T	2,2-Dichloropropane	0.855	0.760	11.1	86	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.604	2.9	94	0.00
28 T	Bromochloromethane	0.468	0.400	14.5	84	0.00
29 T	Tetrahydrofuran	0.254	0.267	-5.1	97	0.00
30 C	Chloroform	0.987	1.021	-3.4#	99	0.00
31 T	Cyclohexane	0.832	0.778	6.5	88	0.00
32 T	1,1,1-Trichloroethane	0.909	0.949	-4.4	101	0.00
33 S	1,2-Dichloroethane-d4	0.616	0.593	3.7	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.316	0.325	-2.8	98	0.00
36 T	1,1-Dichloropropene	0.473	0.477	-0.8	94	0.00
37 T	Ethyl Acetate	0.479	0.527	-10.0	97	0.00
38 T	Carbon Tetrachloride	0.520	0.571	-9.8	101	0.00
39 T	Methylcyclohexane	0.526	0.516	1.9	88	0.00
40 TM	Benzene	1.388	1.403	-1.1	93	0.00
41 T	Methacrylonitrile	0.270	0.303	-12.2	103	0.00
42 TM	1,2-Dichloroethane	0.502	0.552	-10.0	102	0.00
43 T	Isopropyl Acetate	0.797	0.852	-6.9	97	0.00
44 TM	Trichloroethene	0.418	0.413	1.2	95	0.00
45 C	1,2-Dichloropropane	0.357	0.371	-3.9#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX070820\
 Data File : VX017273.D
 Acq On : 08 Jul 2020 22:16
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 05:39:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.243	0.263	-8.2	100	0.00
47 T	Bromodichloromethane	0.502	0.558	-11.2	101	0.00
48 T	Methyl methacrylate	0.379	0.428	-12.9	98	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	90	0.00
50 S	Toluene-d8	1.170	1.169	0.1	97	0.00
51 T	4-Methyl-2-Pentanone	0.475	0.547	-15.2	102	0.00
52 CM	Toluene	0.867	0.924	-6.6#	97	0.00
53 T	t-1,3-Dichloropropene	0.554	0.610	-10.1	100	0.00
54 T	cis-1,3-Dichloropropene	0.593	0.629	-6.1	96	0.00
55 T	1,1,2-Trichloroethane	0.354	0.401	-13.3	105	0.00
56 T	Ethyl methacrylate	0.504	0.594	-17.9	102	0.00
57 T	1,3-Dichloropropane	0.599	0.650	-8.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.251	0.273	-8.8	91	0.00
59 T	2-Hexanone	0.347	0.423	-21.9#	106	0.00
60 T	Dibromochloromethane	0.411	0.480	-16.8	106	0.00
61 T	1,2-Dibromoethane	0.381	0.424	-11.3	102	0.00
62 S	4-Bromofluorobenzene	0.457	0.483	-5.7	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.404	0.409	-1.2	97	0.00
65 PM	Chlorobenzene	1.032	1.074	-4.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.435	-9.3	104	0.00
67 C	Ethyl Benzene	1.723	1.866	-8.3#	99	0.00
68 T	m/p-Xylenes	0.662	0.720	-8.8	99	0.00
69 T	o-Xylene	0.630	0.698	-10.8	102	0.00
70 T	Styrene	1.072	1.250	-16.6	103	0.00
71 P	Bromoform	0.315	0.391	-24.1#	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.275	3.581	-9.3	103	0.00
74 T	N-amyl acetate	1.397	1.549	-10.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.030	1.087	-5.5	104	0.00
76 T	1,2,3-Trichloropropane	0.983	1.073	-9.2	103	0.00
77 T	Bromobenzene	0.900	0.984	-9.3	107	0.00
78 T	n-propylbenzene	3.704	4.058	-9.6	103	0.00
79 T	2-Chlorotoluene	2.241	2.462	-9.9	104	0.00
80 T	1,3,5-Trimethylbenzene	2.717	3.038	-11.8	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.380	0.391	-2.9	95	0.00
82 T	4-Chlorotoluene	2.679	2.926	-9.2	104	0.00
83 T	tert-Butylbenzene	2.667	2.987	-12.0	104	0.00
84 T	1,2,4-Trimethylbenzene	2.756	3.114	-13.0	104	0.00
85 T	sec-Butylbenzene	3.135	3.436	-9.6	102	0.00
86 T	p-Isopropyltoluene	2.918	3.239	-11.0	101	0.00
87 T	1,3-Dichlorobenzene	1.635	1.688	-3.2	105	0.00
88 T	1,4-Dichlorobenzene	1.675	1.718	-2.6	104	0.00
89 T	n-Butylbenzene	2.606	2.638	-1.2	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070820\
Data File : VX017273.D
Acq On : 08 Jul 2020 22:16
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 05:39:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.575	0.613	-6.6	102	0.00
91 T	1,2-Dichlorobenzene	1.528	1.622	-6.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.246	0.273	-11.0	112	0.00
93 T	1,2,4-Trichlorobenzene	1.043	1.128	-8.1	106	0.00
94 T	Hexachlorobutadiene	0.422	0.413	2.1	101	0.00
95 T	Naphthalene	3.171	3.628	-14.4	104	0.00
96 T	1,2,3-Trichlorobenzene	1.009	1.097	-8.7	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6