

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070820\
 Data File : VX017275.D
 Acq On : 08 Jul 2020 23:02
 Operator : JC/SP
 Sample : VX0708WBL03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0708WBL03

Quant Time: Jul 09 05:42:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241977	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	401822	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	416248	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220430	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	159163	53.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.72%	
35) Dibromofluoromethane	5.47	113	136382	53.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.32%	
50) Toluene-d8	8.70	98	478585	50.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.84%	
62) 4-Bromofluorobenzene	11.12	95	206531	56.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.42%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070820\
Data File : VX017275.D
Acq On : 08 Jul 2020 23:02
Operator : JC/SP
Sample : VX0708WBL03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0708WBL03

Quant Time: Jul 09 05:42:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.01 min

Lab File: VX017275.D

Acq: 08 Jul 2020 23:02

Instrument :

MSVOA_X

ClientSampleId :


VX0708WBL03

Tot Ion:168 Resp: 241977

Ion Ratio Lower Upper




168 100

99 57.3 42.6 64.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 53.355 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX017275.D


Acq: 08 Jul 2020 23:02

Tgt Ion: 65 Resp: 159163

Ion Ratio Lower Upper



65 100


67 50.8 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX017275.D

Acq: 08 Jul 2020 23:02

Instrument :

MSVOA_X

ClientSampleId :

VX0708WBL03


Tot Ion:114 Resp: 401822

Ion Ratio Lower Upper

114 100

63 20.6 0.0 41.4




88 15.6 0.0 32.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 53.656 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.01 min

Lab File: VX017275.D

Acq: 08 Jul 2020 23:02


Tgt Ion:113 Resp: 136382

Ion Ratio Lower Upper

113 100

111 102.3 81.9 122.9



192 20.9 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.918 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX017275.D

Acq: 08 Jul 2020 23:02

Instrument :

MSVOA_X

ClientSampleId :


VX0708WBL03

Tot Ion: 98 Resp: 478585

Ion Ratio Lower Upper




98 100

100 67.7 53.8 80.6



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 56.211 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX017275.D

Acq: 08 Jul 2020 23:02


Tgt Ion: 95 Resp: 206531

Ion Ratio Lower Upper

95 100

174 84.8 0.0 164.0



176 81.9 0.0 157.6




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX017275.D
Acq: 08 Jul 2020 23:02

Instrument :
MSVOA_X
ClientSampleId :
VX0708WBL03

Tot Ion:117 Resp: 416248
Ion Ratio Lower Upper
117 100
82 53.2 42.6 63.8
119 32.4 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX017275.D
Acq: 08 Jul 2020 23:02

Tgt Ion:152 Resp: 220430
Ion Ratio Lower Upper
152 100
115 58.7 40.9 122.7
150 160.0 0.0 351.6

