

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023196.D
 Acq On : 08 Jul 2021 13:55
 Operator : JC/MD
 Sample : M2953-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-02-070621

Quant Time: Jul 09 02:10:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	229662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	376761	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	157922	51.163	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.320%
35) Dibromofluoromethane	5.397	113	118843	48.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	8.653	98	450969	49.691	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.085	95	164455	46.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.880%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14887	12.191	ug/l	90
20) Methylene Chloride	2.794	84	12939	4.921	ug/l	91
25) 2-Butanone	4.574	43	4124	2.240	ug/l	93
29) Tetrahydrofuran	5.032	42	2368	2.000	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023196.D
Acq On : 08 Jul 2021 13:55
Operator : JC/MD
Sample : M2953-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-02-070621

Quant Time: Jul 09 02:10:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

