

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023199.D
 Acq On : 08 Jul 2021 15:05
 Operator : JC/MD
 Sample : M2953-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR3-070621

Quant Time: Jul 09 02:11:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	217068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	363360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	325959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	149999	51.415	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.840%	
35) Dibromofluoromethane	5.397	113	113829	48.512	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.020%	
50) Toluene-d8	8.653	98	428533	48.960	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.920%	
62) 4-Bromofluorobenzene	11.085	95	157250	46.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.080%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.373	62	733	0.356	ug/l #	86
9) 1,1,2-Trichlorotrifluo...	2.325	101	1226	0.585	ug/l	97
12) 1,1-Dichloroethene	2.325	96	67064	34.659	ug/l	93
16) Acetone	2.392	43	5842	5.061	ug/l #	88
17) Carbon Disulfide	2.514	76	21185	5.191	ug/l	97
20) Methylene Chloride	2.794	84	1960	0.789	ug/l #	88
21) trans-1,2-Dichloroethene	3.099	96	4374	2.093	ug/l	97
24) 1,1-Dichloroethane	3.617	63	18174	4.538	ug/l	98
27) cis-1,2-Dichloroethene	4.501	96	66411	26.237	ug/l	92
30) Chloroform	5.117	83	3717	0.859	ug/l #	82
42) 1,2-Dichloroethane	6.104	62	13363	3.536	ug/l	95
44) Trichloroethene	7.135	130	370342	138.813	ug/l	100
49) 1,4-Dioxane	7.671	88	1311	20.197	ug/l #	80
52) Toluene	8.726	92	4685	0.770	ug/l	98
64) Tetrachloroethene	9.281	164	2134	0.821	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023199.D
Acq On : 08 Jul 2021 15:05
Operator : JC/MD
Sample : M2953-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR3-070621

Quant Time: Jul 09 02:11:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

