

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023200.D
 Acq On : 08 Jul 2021 15:29
 Operator : JC/MD
 Sample : M2772-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-11

Quant Time: Jul 09 02:11:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	220476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	360773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153820	51.910	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.820%
35) Dibromofluoromethane	5.397	113	112876	48.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.653	98	431432	49.645	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.280%
62) 4-Bromofluorobenzene	11.085	95	153002	45.604	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	22378	19.088	ug/l	# 86
20) Methylene Chloride	2.794	84	14477	5.736	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023200.D
Acq On : 08 Jul 2021 15:29
Operator : JC/MD
Sample : M2772-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-11

Quant Time: Jul 09 02:11:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

