

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023202.D
 Acq On : 08 Jul 2021 16:16
 Operator : JC/MD
 Sample : M2953-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-1.5-6.5-070621

Quant Time: Jul 09 02:12:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	227995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	376064	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	346320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	153638	50.139	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.280%	
35) Dibromofluoromethane	5.397	113	120825	49.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.500%	
50) Toluene-d8	8.653	98	446171	49.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.500%	
62) 4-Bromofluorobenzene	11.085	95	171615	49.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.140%	
Target Compounds						Qvalue
4) Vinyl Chloride	1.374	62	17447	8.058	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.337	101	16707050	7587.101	ug/l	88
12) 1,1-Dichloroethene	2.325	96	207217	101.958	ug/l #	79
20) Methylene Chloride	2.794	84	38452	14.732	ug/l	93
21) trans-1,2-Dichloroethene	3.099	96	9791	4.461	ug/l	90
24) 1,1-Dichloroethane	3.623	63	23304	5.539	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	2811653	1057.562	ug/l	92
29) Tetrahydrofuran	5.025	42	19824	16.870	ug/l	93
30) Chloroform	5.111	83	29612	6.516	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	99269	25.081	ug/l	98
40) Benzene	6.050	78	53910	5.539	ug/l	97
42) 1,2-Dichloroethane	6.098	62	16919	4.326	ug/l	94
44) Trichloroethene	7.202	130	49750351	18017.705	ug/l	94
52) Toluene	8.720	92	6160	0.978	ug/l	95
55) 1,1,2-Trichloroethane	9.159	97	31033	11.732	ug/l	97
64) Tetrachloroethene	9.281	164	726257	262.925	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.219	83	938	0.268	ug/l #	94
88) 1,4-Dichlorobenzene	12.042	146	18037	3.462	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023202.D
Acq On : 08 Jul 2021 16:16
Operator : JC/MD
Sample : M2953-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-1.5-6.5-070621

Quant Time: Jul 09 02:12:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

