

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023205.D
 Acq On : 08 Jul 2021 17:26
 Operator : JC/MD
 Sample : PB137563ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137563ZHE#11

Quant Time: Jul 09 08:13:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	227940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	374862	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	341020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156170	50.977	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.960%		
35) Dibromofluoromethane	5.397	113	117815	48.670	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.340%		
50) Toluene-d8	8.653	98	447574	49.566	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.140%		
62) 4-Bromofluorobenzene	11.085	95	168674	48.386	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.780%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	65297	29.660	ug/l	98
16) Acetone	2.392	43	7658	6.318	ug/l	99
20) Methylene Chloride	2.794	84	7203	2.760	ug/l #	90
27) cis-1,2-Dichloroethene	4.501	96	5309	1.997	ug/l	95
44) Trichloroethene	7.135	130	362779	131.806	ug/l	99
64) Tetrachloroethene	9.275	164	3592	1.321	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023205.D
Acq On : 08 Jul 2021 17:26
Operator : JC/MD
Sample : PB137563ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137563ZHE#11

Quant Time: Jul 09 08:13:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

