

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023206.D
 Acq On : 08 Jul 2021 17:50
 Operator : JC/MD
 Sample : PB137563ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137563ZHE#12

Quant Time: Jul 09 08:13:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	215631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357789	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	311415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150405	51.898	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.800%
35) Dibromofluoromethane	5.397	113	113131	48.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	8.653	98	419897	48.720	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.440%
62) 4-Bromofluorobenzene	11.085	95	145855	43.837	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	32202	15.462	ug/l	96
16) Acetone	2.386	43	7490	6.533	ug/l	92
20) Methylene Chloride	2.794	84	7836	3.174	ug/l	88
44) Trichloroethene	7.135	130	127639	48.587	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023206.D
Acq On : 08 Jul 2021 17:50
Operator : JC/MD
Sample : PB137563ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137563ZHE#12

Quant Time: Jul 09 08:13:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

