

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023210.D
 Acq On : 08 Jul 2021 19:24
 Operator : JC/MD
 Sample : M2772-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1

Quant Time: Jul 09 08:15:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	212597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	349284	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	146525	51.281	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.560%
35) Dibromofluoromethane	5.397	113	106839	47.368	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	8.653	98	411575	48.917	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.085	95	148195	45.625	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	10775	5.248	ug/l	97
16) Acetone	2.386	43	21510	19.028	ug/l	93
20) Methylene Chloride	2.794	84	10612	4.360	ug/l	95
44) Trichloroethene	7.135	130	36993	14.425	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023210.D
Acq On : 08 Jul 2021 19:24
Operator : JC/MD
Sample : M2772-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1

Quant Time: Jul 09 08:15:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

