

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023211.D
 Acq On : 08 Jul 2021 19:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 08:15:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	52.687	-5.4	93	0.00
3 P	Chloromethane	50.000	49.019	2.0	93	0.00
4 C	Vinyl Chloride	50.000	50.553	-1.1#	93	0.00
5 T	Bromomethane	50.000	44.227	11.5	83	0.00
6 T	Chloroethane	50.000	48.609	2.8	89	0.00
7 T	Trichlorofluoromethane	50.000	49.685	0.6	90	0.00
8 T	Diethyl Ether	50.000	50.118	-0.2	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	55.464	-10.9	104	0.00
10 T	Methyl Iodide	50.000	48.318	3.4	90	0.00
11 T	Tert butyl alcohol	250.000	255.888	-2.4	102	0.00
12 CM	1,1-Dichloroethene	50.000	50.500	-1.0#	93	0.00
13 T	Acrolein	250.000	174.394	30.2#	67	0.00
14 T	Allyl chloride	50.000	52.728	-5.5	97	0.00
15 T	Acrylonitrile	250.000	260.550	-4.2	96	0.00
16 T	Acetone	250.000	245.802	1.7	97	0.00
17 T	Carbon Disulfide	50.000	47.469	5.1	91	0.00
18 T	Methyl Acetate	50.000	52.106	-4.2	98	0.00
19 T	Methyl tert-butyl Ether	50.000	53.857	-7.7	100	0.00
20 T	Methylene Chloride	50.000	47.752	4.5	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.019	-2.0	96	0.00
22 T	Diisopropyl ether	50.000	52.053	-4.1	96	0.00
23 T	Vinyl Acetate	250.000	265.659	-6.3	95	0.00
24 P	1,1-Dichloroethane	50.000	52.225	-4.5	97	0.00
25 T	2-Butanone	250.000	255.356	-2.1	95	0.00
26 T	2,2-Dichloropropane	50.000	49.379	1.2	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.643	-3.3	97	0.00
28 T	Bromochloromethane	50.000	47.558	4.9	91	0.00
29 T	Tetrahydrofuran	250.000	253.659	-1.5	94	0.00
30 C	Chloroform	50.000	52.024	-4.0#	97	0.00
31 T	Cyclohexane	50.000	48.283	3.4	93	0.00
32 T	1,1,1-Trichloroethane	50.000	53.753	-7.5	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.505	1.0	97	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	48.716	2.6	99	0.00
36 T	1,1-Dichloropropene	50.000	49.475	1.0	95	0.00
37 T	Ethyl Acetate	50.000	50.922	-1.8	96	0.00
38 T	Carbon Tetrachloride	50.000	53.215	-6.4	99	0.00
39 T	Methylcyclohexane	50.000	49.467	1.1	96	0.00
40 TM	Benzene	50.000	50.009	-0.0	95	0.00
41 T	Methacrylonitrile	50.000	52.179	-4.4	100	0.00
42 TM	1,2-Dichloroethane	50.000	52.634	-5.3	98	0.00
43 T	Isopropyl Acetate	50.000	50.074	-0.1	95	0.00
44 TM	Trichloroethene	50.000	59.263	-18.5	113	0.00
45 C	1,2-Dichloropropane	50.000	51.093	-2.2#	97	0.00
46 T	Dibromomethane	50.000	51.444	-2.9	97	0.00
47 T	Bromodichloromethane	50.000	54.590	-9.2	100	0.00
48 T	Methyl methacrylate	50.000	50.882	-1.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
 Data File : VX023211.D
 Acq On : 08 Jul 2021 19:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 08:15:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1018.639	-1.9	95	0.00
50 S	Toluene-d8	50.000	48.081	3.8	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	255.466	-2.2	95	0.00
52 CM	Toluene	50.000	51.204	-2.4#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	47.239	5.5	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.913	4.2	96	0.00
55 T	1,1,2-Trichloroethane	50.000	51.560	-3.1	97	0.00
56 T	Ethyl methacrylate	50.000	53.070	-6.1	99	0.00
57 T	1,3-Dichloropropane	50.000	51.365	-2.7	97	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	243.062	2.8	94	0.00
59 T	2-Hexanone	250.000	254.746	-1.9	93	0.00
60 T	Dibromochloromethane	50.000	46.837	6.3	97	0.00
61 T	1,2-Dibromoethane	50.000	53.284	-6.6	97	0.00
62 S	4-Bromofluorobenzene	50.000	49.089	1.8	97	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	49.899	0.2	95	0.00
65 PM	Chlorobenzene	50.000	50.951	-1.9	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.627	-11.3	100	0.00
67 C	Ethyl Benzene	50.000	52.125	-4.3#	97	0.00
68 T	m/p-Xylenes	100.000	104.758	-4.8	96	0.00
69 T	o-Xylene	50.000	52.137	-4.3	97	0.00
70 T	Styrene	50.000	53.972	-7.9	97	0.00
71 P	Bromoform	50.000	46.712	6.6	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	52.775	-5.5	96	0.00
74 T	N-ethyl acetate	50.000	50.799	-1.6	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.626	-3.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	53.228	-6.5	96	0.00
77 T	Bromobenzene	50.000	51.451	-2.9	95	0.00
78 T	n-propylbenzene	50.000	51.997	-4.0	94	0.00
79 T	2-Chlorotoluene	50.000	51.868	-3.7	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.517	-5.0	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.066	1.9	96	0.00
82 T	4-Chlorotoluene	50.000	52.451	-4.9	94	0.00
83 T	tert-Butylbenzene	50.000	53.654	-7.3	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.214	-6.4	95	0.00
85 T	sec-Butylbenzene	50.000	53.226	-6.5	95	0.00
86 T	p-Isopropyltoluene	50.000	52.964	-5.9	94	0.00
87 T	1,3-Dichlorobenzene	50.000	51.286	-2.6	94	0.00
88 T	1,4-Dichlorobenzene	50.000	51.520	-3.0	95	0.00
89 T	n-Butylbenzene	50.000	52.180	-4.4	92	0.00
90 T	Hexachloroethane	50.000	48.282	3.4	98	0.00
91 T	1,2-Dichlorobenzene	50.000	51.826	-3.7	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.585	0.8	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.601	-5.2	96	0.00
94 T	Hexachlorobutadiene	50.000	52.693	-5.4	98	0.00
95 T	Naphthalene	50.000	49.660	0.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070821\
Data File : VX023211.D
Acq On : 08 Jul 2021 19:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 08:15:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.458	-6.9	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6