

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003260.D
 Acq On : 09 Jul 2018 22:44
 Operator : JC/MD
 Sample : J3762-08 LOO 10PPB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOQ-WATER-02-QT3-2018

Manual Integrations
 APPROVED

MMDadoda
 7/10/2018 3:11:54 PM

Quant Time: Jul 10 11:22:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	309480	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	291193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	203537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193026	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
35) Dibromofluoromethane	5.51	113	160535	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	8.72	98	590568	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	11.14	95	220604	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
Target Compounds						
10) Methyl Iodide	2.51	142	25305	12.43	ug/l	97
11) Tert butyl alcohol	3.09	59	38972	56.26	ug/l	97
13) Acrolein	2.29	56	30768	55.12	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003260.D
Acq On : 09 Jul 2018 22:44
Operator : JC/MD
Sample : J3762-08 LOO 10PPB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LOQ-WATER-02-QT3-2018

Manual Integrations
APPROVED

MMDadoda
7/10/2018 3:11:54 PM

Quant Time: Jul 10 11:22:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

