

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003264.D
 Acq On : 10 Jul 2018 00:30
 Operator : JC/MD
 Sample : PB110911ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110911ZHE#10

Quant Time: Jul 10 01:28:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267399	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	368785	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	345749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	194060	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	169280	40.23	ug/l	0.00
Spiked Amount			Recovery	=	80.46%	
35) Dibromofluoromethane	5.51	113	144762	36.54	ug/l	0.00
Spiked Amount			Recovery	=	73.08%	
50) Toluene-d8	8.72	98	513225	36.47	ug/l	0.00
Spiked Amount			Recovery	=	72.94%	
62) 4-Bromofluorobenzene	11.14	95	179066	34.44	ug/l	0.00
Spiked Amount			Recovery	=	68.88%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	3731	6.56	ug/l	95
16) Acetone	2.45	43	23179	15.51	ug/l	99
18) Methyl Acetate	2.78	43	22099	5.24	ug/l	94
20) Methylene Chloride	2.86	84	5185	1.00	ug/l #	82
25) 2-Butanone	4.72	43	5348	2.02	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003264.D
Acq On : 10 Jul 2018 00:30
Operator : JC/MD
Sample : PB110911ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110911ZHE#10

Quant Time: Jul 10 01:28:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

