

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003266.D
 Acq On : 10 Jul 2018 01:24
 Operator : JC/MD
 Sample : PB110911ZHE#12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110911ZHE#12

Quant Time: Jul 10 01:49:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274297	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	381127	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	204160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	175263	40.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.20%	
35) Dibromofluoromethane	5.51	113	149507	36.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.02%	
50) Toluene-d8	8.72	98	537983	36.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.98%	
62) 4-Bromofluorobenzene	11.14	95	186400	34.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.38%	
Target Compounds						
10) Methyl Iodide	2.52	142	2797	6.31	ug/l	90
16) Acetone	2.45	43	24166	15.77	ug/l	100
18) Methyl Acetate	2.78	43	23443	5.42	ug/l	95
20) Methylene Chloride	2.86	84	5349	1.01	ug/l	87
25) 2-Butanone	4.71	43	5755	2.12	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003266.D
Acq On : 10 Jul 2018 01:24
Operator : JC/MD
Sample : PB110911ZHE#12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110911ZHE#12

Quant Time: Jul 10 01:49:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

