

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003267.D
 Acq On : 10 Jul 2018 01:50
 Operator : JC/MD
 Sample : PB110911ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110911ZHE#13

Quant Time: Jul 10 03:27:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	258212	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362830	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	345205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195741	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168409	41.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.88%	
35) Dibromofluoromethane	5.51	113	140194	35.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.94%	
50) Toluene-d8	8.72	98	519780	37.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.08%	
62) 4-Bromofluorobenzene	11.14	95	179385	35.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.12%	
Target Compounds						
10) Methyl Iodide	2.52	142	2259	6.22	ug/l	95
16) Acetone	2.45	43	24002	16.64	ug/l	98
18) Methyl Acetate	2.78	43	24175	5.94	ug/l	97
20) Methylene Chloride	2.86	84	5057	1.02	ug/l	86
25) 2-Butanone	4.71	43	5073	1.99	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003267.D
Acq On : 10 Jul 2018 01:50
Operator : JC/MD
Sample : PB110911ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110911ZHE#13

Quant Time: Jul 10 03:27:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

