

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003268.D
 Acq On : 10 Jul 2018 02:17
 Operator : JC/MD
 Sample : PB110911ZHE#14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110911ZHE#14

Quant Time: Jul 10 03:30:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	370963	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	346868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196432	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	164277	38.36	ug/l	0.00
Spiked Amount			Recovery	=	76.72%	
35) Dibromofluoromethane	5.51	113	143739	36.07	ug/l	0.00
Spiked Amount			Recovery	=	72.14%	
50) Toluene-d8	8.72	98	515224	36.40	ug/l	0.00
Spiked Amount			Recovery	=	72.80%	
62) 4-Bromofluorobenzene	11.14	95	180263	34.46	ug/l	0.00
Spiked Amount			Recovery	=	68.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1790	6.08	ug/l	# 91
16) Acetone	2.45	43	24854	16.35	ug/l	98
18) Methyl Acetate	2.78	43	22609	5.27	ug/l	92
20) Methylene Chloride	2.86	84	4937	0.89	ug/l	91
25) 2-Butanone	4.71	43	5221	1.94	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003268.D
Acq On : 10 Jul 2018 02:17
Operator : JC/MD
Sample : PB110911ZHE#14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110911ZHE#14

Quant Time: Jul 10 03:30:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

