

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003270.D
 Acq On : 10 Jul 2018 03:10
 Operator : JC/MD
 Sample : J3585-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01RE

Quant Time: Jul 10 11:37:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	271840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	373048	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	185592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166782	38.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.96%	
35) Dibromofluoromethane	5.51	113	144395	36.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.06%	
50) Toluene-d8	8.72	98	510958	35.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.78%	
62) 4-Bromofluorobenzene	11.14	95	171666	32.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.28%	
Target Compounds						
4) Vinyl Chloride	1.40	62	13542	3.51	ug/l	# 64
16) Acetone	2.45	43	13702	9.02	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	192185	48.02	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003270.D
Acq On : 10 Jul 2018 03:10
Operator : JC/MD
Sample : J3585-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01RE

Quant Time: Jul 10 11:37:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

