

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003272.D
 Acq On : 10 Jul 2018 04:04
 Operator : JC/MD
 Sample : J3896-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-SOIL

Quant Time: Jul 10 05:20:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358519	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	335398	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	184847	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	165148	41.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.10%	
35) Dibromofluoromethane	5.51	113	139031	36.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.20%	
50) Toluene-d8	8.72	98	501475	36.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.30%	
62) 4-Bromofluorobenzene	11.14	95	176858	34.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1625	6.07	ug/l	# 96
16) Acetone	2.45	43	21798	15.45	ug/l	96
18) Methyl Acetate	2.78	43	22259	5.59	ug/l	94
20) Methylene Chloride	2.86	84	10882	3.10	ug/l	93
25) 2-Butanone	4.72	43	5590	2.24	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003272.D
Acq On : 10 Jul 2018 04:04
Operator : JC/MD
Sample : J3896-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-SOIL

Quant Time: Jul 10 05:20:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

