

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070918\
 Data File : VX003276.D
 Acq On : 10 Jul 2018 05:50
 Operator : JC/MD
 Sample : J3891-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 212-72-11940-R4-VNJ-242

Quant Time: Jul 10 06:41:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	345412	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	321961	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184949	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	158947	40.32	ug/l	0.00
Spiked Amount			Recovery	=	80.64%	
35) Dibromofluoromethane	5.51	113	135508	36.52	ug/l	0.00
Spiked Amount			Recovery	=	73.04%	
50) Toluene-d8	8.72	98	486495	36.91	ug/l	0.00
Spiked Amount			Recovery	=	73.82%	
62) 4-Bromofluorobenzene	11.14	95	169067	34.71	ug/l	0.00
Spiked Amount			Recovery	=	69.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1291	5.99	ug/l	98
16) Acetone	2.45	43	26553	18.97	ug/l	94
18) Methyl Acetate	2.78	43	22315	5.65	ug/l	96
20) Methylene Chloride	2.86	84	483147	170.17	ug/l	93
25) 2-Butanone	4.72	43	6395	2.58	ug/l	92
95) Naphthalene	13.84	128	333456	22.55	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070918\
Data File : VX003276.D
Acq On : 10 Jul 2018 05:50
Operator : JC/MD
Sample : J3891-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
212-72-11940-R4-VNJ-242

Quant Time: Jul 10 06:41:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

