

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070919\
 Data File : VX010744.D
 Acq On : 09 Jul 2019 12:40
 Operator : JC/SP
 Sample : K3692-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 10 02:21:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	217827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	341204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	315213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146618	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	109486	53.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.22%	
35) Dibromofluoromethane	5.50	113	105523	50.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
50) Toluene-d8	8.71	98	404223	50.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.13	95	138112	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070919\
Data File : VX010744.D
Acq On : 09 Jul 2019 12:40
Operator : JC/SP
Sample : K3692-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 10 02:21:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

