

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070920\
 Data File : VX017309.D
 Acq On : 09 Jul 2020 15:48
 Operator : JC/SP
 Sample : L3235-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 10 06:50:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X070920W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 09 12:25:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.99	128	51082	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.83	114	253941	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	241726	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	111020	30.24	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.80%
60) 4-Bromofluorobenzene	11.12	95	106960	25.94	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	86.47%
63) Toluene-d8	8.70	98	301968	29.53	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.43%

Target Compounds

					Ovalue	
55) 2-Chloroethyl vinyl ether	8.29	63	2206	19.636	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070920\
Data File : VX017309.D
Acq On : 09 Jul 2020 15:48
Operator : JC/SP
Sample : L3235-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 10 06:50:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X070920W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 09 12:25:06 2020
Response via : Initial Calibration

