

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070920\
 Data File : VX017317.D
 Acq On : 09 Jul 2020 18:50
 Operator : JC/SP
 Sample : L3217-11 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E3-41-NHS

Quant Time: Jul 10 07:27:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	262847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	432106	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	249521	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	169335	52.26	ug/l	0.00
Spiked Amount			Recovery	=	104.52%	
35) Dibromofluoromethane	5.46	113	144956	53.03	ug/l	0.00
Spiked Amount			Recovery	=	106.06%	
50) Toluene-d8	8.70	98	531903	52.62	ug/l	0.00
Spiked Amount			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	11.12	95	229951	58.20	ug/l	0.00
Spiked Amount			Recovery	=	116.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	8376	6.910	ug/l	89
20) Methylene Chloride	2.84	84	4933	1.007	ug/l #	90
30) Chloroform	5.17	83	10198	1.966	ug/l	91
32) 1,1,1-Trichloroethane	5.47	97	9992	2.091	ug/l	98
52) Toluene	8.77	92	8965	1.197	ug/l	97
68) m/p-Xylenes	10.34	106	12005	1.985	ug/l	98
69) o-Xylene	10.68	106	12883	2.239	ug/l	98
78) n-propylbenzene	11.34	91	24638	1.333	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	72871	5.375	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	294591	21.419	ug/l	98
89) n-Butylbenzene	12.37	91	18290	1.407	ug/l #	35
91) 1,2-Dichlorobenzene	12.38	146	10738	1.408	ug/l	93
95) Naphthalene	13.82	128	42884	2.710	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070920\
Data File : VX017317.D
Acq On : 09 Jul 2020 18:50
Operator : JC/SP
Sample : L3217-11 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
E3-41-NHS

Quant Time: Jul 10 07:27:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

