

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070920\
 Data File : VX017320.D
 Acq On : 09 Jul 2020 19:58
 Operator : JC/SP
 Sample : L3217-14 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E3-41-BOILER-ROOM

Quant Time: Jul 10 07:32:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	243923	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	409377	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	426489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226259	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	159277	52.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.94%	
35) Dibromofluoromethane	5.47	113	137792	53.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.42%	
50) Toluene-d8	8.70	98	499051	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.12	95	205217	54.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.64%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070920\
Data File : VX017320.D
Acq On : 09 Jul 2020 19:58
Operator : JC/SP
Sample : L3217-14 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E3-41-BOILER-ROOM

Quant Time: Jul 10 07:32:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.01 min

Lab File: VX017320.D

Acq: 09 Jul 2020 19:58

Instrument :

MSVOA_X

ClientSampleId :


E3-41-BOILER-ROOM

Tot Ion:168 Resp: 243923

Ion Ratio Lower Upper




168 100

99 58.1 42.6 64.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 52.967 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX017320.D


Acq: 09 Jul 2020 19:58

Tgt Ion: 65 Resp: 159277

Ion Ratio Lower Upper



65 100


67 52.1 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX017320.D

Acq: 09 Jul 2020 19:58

Instrument :

MSVOA_X

ClientSampleId :

E3-41-BOILER-ROOM


Tot Ion:114 Resp: 409377

Ion Ratio Lower Upper

114 100

63 19.3 0.0 41.4




88 15.4 0.0 32.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 53.210 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.01 min

Lab File: VX017320.D

Acq: 09 Jul 2020 19:58


Tgt Ion:113 Resp: 137792

Ion Ratio Lower Upper

113 100

111 100.9 81.9 122.9



192 21.9 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.116 ug/l
RT: 8.70 min Scan# 1248
Delta R.T. -0.00 min
Lab File: VX017320.D
Acq: 09 Jul 2020 19:58




Instrument :
MSVOA_X
ClientSampleId :
E3-41-BOILER-ROOM


Tot Ion: 98 Resp: 499051
Ion Ratio Lower Upper
98 100
100 66.3 53.8 80.6



#62
4-Bromofluorobenzene
Concen: 54.823 ug/l
RT: 11.12 min Scan# 1646
Delta R.T. -0.00 min
Lab File: VX017320.D
Acq: 09 Jul 2020 19:58

Tgt Ion: 95 Resp: 205217
Ion Ratio Lower Upper
95 100
174 84.9 0.0 164.0
176 82.8 0.0 157.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX017320.D
Acq: 09 Jul 2020 19:58

Instrument :
MSVOA_X
ClientSampleId :
E3-41-BOILER-ROOM

Tot Ion:117 Resp: 426489
Ion Ratio Lower Upper
117 100
82 53.7 42.6 63.8
119 31.3 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX017320.D
Acq: 09 Jul 2020 19:58

Tgt Ion:152 Resp: 226259
Ion Ratio Lower Upper
152 100
115 58.2 40.9 122.7
150 157.5 0.0 351.6

