

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
 Data File : VX023221.D
 Acq On : 09 Jul 2021 15:15
 Operator : JC/MD
 Sample : PB137563TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137563TB

Quant Time: Jul 10 04:20:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	208548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	343695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	310805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	141616	50.525	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.060%
35) Dibromofluoromethane	5.397	113	110765	49.907	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	411671	49.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.085	95	148357	46.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	10454	9.427	ug/l #	85
20) Methylene Chloride	2.800	84	4610	1.931	ug/l	94
44) Trichloroethene	7.135	130	5490	2.176	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
Data File : VX023221.D
Acq On : 09 Jul 2021 15:15
Operator : JC/MD
Sample : PB137563TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137563TB

Quant Time: Jul 10 04:20:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

