

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
 Data File : VX023223.D
 Acq On : 09 Jul 2021 16:02
 Operator : JC/MD
 Sample : M2953-10DL2 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-6.5-11.5-070621DL2

Quant Time: Jul 10 04:20:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	209714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	344850	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	307275	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	144859	51.395	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.780%	
35) Dibromofluoromethane	5.397	113	109671	49.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.500%	
50) Toluene-d8	8.653	98	411667	49.558	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.120%	
62) 4-Bromofluorobenzene	11.085	95	147824	46.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.200%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	19124	9.442	ug/l	99
20) Methylene Chloride	2.794	84	2856	1.190	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	2543	1.040	ug/l	96
44) Trichloroethene	7.135	130	124487	49.165	ug/l	99
64) Tetrachloroethene	9.275	164	704	0.287	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
Data File : VX023223.D
Acq On : 09 Jul 2021 16:02
Operator : JC/MD
Sample : M2953-10DL2 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-6.5-11.5-070621DL2

Quant Time: Jul 10 04:20:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

