

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
 Data File : VX023224.D
 Acq On : 09 Jul 2021 16:25
 Operator : JC/MD
 Sample : PB137563ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137563ZHE#12

Quant Time: Jul 10 04:21:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	212760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	348774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143959	50.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.680%	
35) Dibromofluoromethane	5.397	113	112021	49.738	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.480%	
50) Toluene-d8	8.653	98	412144	49.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.120%	
62) 4-Bromofluorobenzene	11.085	95	148176	45.685	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.380%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6819	6.028	ug/l	98
20) Methylene Chloride	2.794	84	7954	3.266	ug/l	94
44) Trichloroethene	7.135	130	3093	1.208	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
Data File : VX023224.D
Acq On : 09 Jul 2021 16:25
Operator : JC/MD
Sample : PB137563ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137563ZHE#12

Quant Time: Jul 10 04:21:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

