

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
 Data File : VX023227.D
 Acq On : 09 Jul 2021 17:36
 Operator : JC/MD
 Sample : M2772-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1

Quant Time: Jul 10 04:22:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	208941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	342894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	303794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143635	51.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.300%
35) Dibromofluoromethane	5.397	113	107102	48.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	8.653	98	404716	48.999	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.085	95	142228	44.604	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.200%
Target Compounds						
16) Acetone	2.386	43	20361	18.327	ug/l	95
20) Methylene Chloride	2.794	84	9523	3.981	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
Data File : VX023227.D
Acq On : 09 Jul 2021 17:36
Operator : JC/MD
Sample : M2772-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1

Quant Time: Jul 10 04:22:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

