

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
 Data File : VX023228.D
 Acq On : 09 Jul 2021 18:00
 Operator : JC/MD
 Sample : M2748-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-8A

Quant Time: Jul 10 04:22:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	209788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	339801	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	297567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	142502	50.541	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.080%			
35) Dibromofluoromethane	5.397	113	109496	49.900	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.800%			
50) Toluene-d8	8.653	98	398718	48.712	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.420%			
62) 4-Bromofluorobenzene	11.085	95	141265	44.705	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.400%			
Target Compounds						
16) Acetone	2.385	43	27104	24.298	ug/l	92
20) Methylene Chloride	2.794	84	12251	5.101	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070921\
Data File : VX023228.D
Acq On : 09 Jul 2021 18:00
Operator : JC/MD
Sample : M2748-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-8A

Quant Time: Jul 10 04:22:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

