

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
 Data File : VX042249.D
 Acq On : 09 Jul 2024 16:02
 Operator : JC/MD
 Sample : P3188-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240708009-01

Quant Time: Jul 10 04:58:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	27307	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	142589	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	130276	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	55696	29.629	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.767%
60) 4-Bromofluorobenzene	11.079	95	54049	26.802	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.333%
63) Toluene-d8	8.647	98	162442	29.145	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.133%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	687	2.378	ug/l	77
53) Dibromochloromethane	9.525	129	1808	1.040	ug/l	97
56) Bromoform	10.799	173	1200	0.847	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
Data File : VX042249.D
Acq On : 09 Jul 2024 16:02
Operator : JC/MD
Sample : P3188-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240708009-01

Quant Time: Jul 10 04:58:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

