

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
 Data File : VX042251.D
 Acq On : 09 Jul 2024 16:49
 Operator : JC/MD
 Sample : P3188-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240708002-03-TRIP-BLANKRE

Quant Time: Jul 10 04:59:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	24338	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	136175	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	121339	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	52728	31.472	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.900%
60) 4-Bromofluorobenzene	11.085	95	50584	26.932	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.767%
63) Toluene-d8	8.653	98	152161	29.311	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.700%
Target Compounds						
25) Chloroform	5.099	83	8822	3.273	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
Data File : VX042251.D
Acq On : 09 Jul 2024 16:49
Operator : JC/MD
Sample : P3188-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240708002-03-TRIP-BLANKRE

Quant Time: Jul 10 04:59:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

