

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
 Data File : VX042252.D
 Acq On : 09 Jul 2024 17:13
 Operator : JC/MD
 Sample : P3150-01DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240628030-01DL

Quant Time: Jul 09 17:43:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	24680	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	135402	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	121286	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	51323	30.209	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.700%
60) 4-Bromofluorobenzene	11.085	95	49867	26.562	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.533%
63) Toluene-d8	8.653	98	151552	29.206	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.367%
Target Compounds						
15) Acetone	2.392	58	29985	114.842	ug/l	Qvalue 93
18) Methylene Chloride	2.788	84	1286	0.743	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070924\
Data File : VX042252.D
Acq On : 09 Jul 2024 17:13
Operator : JC/MD
Sample : P3150-01DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240628030-01DL

Quant Time: Jul 09 17:43:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

