

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070925\  
 Data File : VX046921.D  
 Acq On : 09 Jul 2025 10:15  
 Operator : JC/MD  
 Sample : VX0709MBL01  
 Misc : 5.00g/10mL/100uL/5./MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 VX0709MBL01

Quant Time: Jul 10 02:54:50 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070225W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Jul 10 02:52:52 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	330948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	572860	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	527826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	258108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	221728	50.714	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.420%	
35) Dibromofluoromethane	5.391	113	190171	48.905	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.800%	
50) Toluene-d8	8.647	98	695530	50.695	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.400%	
62) 4-Bromofluorobenzene	11.079	95	260483	49.956	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070925\  
 Data File : VX046921.D  
 Acq On : 09 Jul 2025 10:15  
 Operator : JC/MD  
 Sample : VX0709MBL01  
 Misc : 5.00g/10mL/100uL/5./MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0709MBL01

Quant Time: Jul 10 02:54:50 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070225W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Jul 10 02:52:52 2025  
 Response via : Initial Calibration

