

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
 Data File : VX046922.D
 Acq On : 09 Jul 2025 10:36
 Operator : JC/MD
 Sample : VX0709WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 VX0709WBL01

Quant Time: Jul 10 02:55:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 10 02:52:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.562	168	444130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	779727	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	725701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	358830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	306664	52.266	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.540%
35) Dibromofluoromethane	5.397	113	258884	48.912	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	8.647	98	937474	50.202	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.079	95	356974	50.298	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
 Data File : VX046922.D
 Acq On : 09 Jul 2025 10:36
 Operator : JC/MD
 Sample : VX0709WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0709WBL01

Quant Time: Jul 10 02:55:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 10 02:52:52 2025
 Response via : Initial Calibration

