

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
 Data File : VX046925.D
 Acq On : 09 Jul 2025 11:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 10 02:58:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 10 02:52:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	365508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	639867	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	587758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	287399	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	246452	51.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.080%
35) Dibromofluoromethane	5.397	113	209049	48.130	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	8.647	98	764167	49.865	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.079	95	289034	49.626	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
Data File : VX046925.D
Acq On : 09 Jul 2025 11:49
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 10 02:58:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 10 02:52:52 2025
Response via : Initial Calibration

