

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
 Data File : VX046930.D
 Acq On : 09 Jul 2025 14:27
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 10 03:14:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 10 02:52:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	324866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	563541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	517518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	255473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	218062	50.809	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.620%	
35) Dibromofluoromethane	5.397	113	184593	48.255	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.520%	
50) Toluene-d8	8.653	98	675509	50.050	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.100%	
62) 4-Bromofluorobenzene	11.079	95	253820	49.483	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	98.960%	
Target Compounds						
20) Methylene Chloride	2.800	84	11677	2.962	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070925\
 Data File : VX046930.D
 Acq On : 09 Jul 2025 14:27
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 10 03:14:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 10 02:52:52 2025
 Response via : Initial Calibration

