

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003286.D
 Acq On : 10 Jul 2018 16:08
 Operator : JC/MD
 Sample : PB110944TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110944TB

Quant Time: Jul 10 17:04:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358555	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	334214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	196661	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163460	40.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.02%	
35) Dibromofluoromethane	5.51	113	137781	35.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.54%	
50) Toluene-d8	8.72	98	501702	36.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.34%	
62) 4-Bromofluorobenzene	11.14	95	178795	35.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.72%	
Target Compounds						
16) Acetone	2.45	43	25815	17.80	ug/l	97
18) Methyl Acetate	2.78	43	22392	5.47	ug/l	97
20) Methylene Chloride	2.86	84	11469	3.20	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003286.D
Acq On : 10 Jul 2018 16:08
Operator : JC/MD
Sample : PB110944TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110944TB

Quant Time: Jul 10 17:04:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

