

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003288.D
 Acq On : 10 Jul 2018 17:01
 Operator : JC/MD
 Sample : PB110944ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110944ZHE#02

Quant Time: Jul 10 18:12:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	357441	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336491	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	188176	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166028	40.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.86%	
35) Dibromofluoromethane	5.51	113	139022	36.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.40%	
50) Toluene-d8	8.72	98	504189	36.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.92%	
62) 4-Bromofluorobenzene	11.14	95	176620	35.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.08%	
Target Compounds						
16) Acetone	2.45	43	25487	17.70	ug/l	95
18) Methyl Acetate	2.78	43	21375	5.26	ug/l	96
20) Methylene Chloride	2.86	84	7238	1.77	ug/l	91
25) 2-Butanone	4.71	43	6349	2.49	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003288.D
Acq On : 10 Jul 2018 17:01
Operator : JC/MD
Sample : PB110944ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110944ZHE#02

Quant Time: Jul 10 18:12:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

