

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003289.D
 Acq On : 10 Jul 2018 17:28
 Operator : JC/MD
 Sample : PB110944ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110944ZHE#03

Quant Time: Jul 10 18:15:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	351487	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	192523	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161894	40.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.72%	
35) Dibromofluoromethane	5.51	113	135669	35.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.86%	
50) Toluene-d8	8.72	98	494990	36.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.80%	
62) 4-Bromofluorobenzene	11.14	95	175091	35.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.66%	
Target Compounds						
16) Acetone	2.45	43	26706	18.75	ug/l	99
18) Methyl Acetate	2.78	43	22954	5.71	ug/l	94
20) Methylene Chloride	2.86	84	8032	2.08	ug/l	89
76) 1,2,3-Trichloropropane	11.14	75	86269	20.73	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003289.D
Acq On : 10 Jul 2018 17:28
Operator : JC/MD
Sample : PB110944ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110944ZHE#03

Quant Time: Jul 10 18:15:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

